



this document brought to you by



ZOLL AED Plus Trainer2 Manual

Get an original copy of the ZOLL AED Plus Trainer2 Manual for manufacturer information about service, available accessories and how to use and maintain your device.

(844) 277-2527
www.GoRescue.com

CONTROLS / INDICATORS

Handheld Remote Controller:

The buttons above the line on the handheld Remote Controller are Set-up Controls, which the instructor uses to select a training scenario (an LED light next to the selected scenario). The buttons below the line are Training Controls, which the instructor uses to introduce events into the training scenario.

Set-up Controls:



View/Enter scenario
Show current scenario selection (press once)
Start scenario (press again)



Move Up to next scenario.



CPR Metronome On/Off Button
Enables the CPR metronome to beep for the selected Training Scenario. The LED above the button lights to indicate that the CPR metronome is enabled. You should enable the CPR metronome whenever CPR-D padz® are connected to the Trainer2.



When a CPR period occurs in the Training Scenario, the CPR metronome begins to beep at 60 compressions per minute (CPM), then automatically accelerates to 100 CPM after a few seconds. Metronome beeps are independent of the student's actual compression rate and continue until either the CPR interval ends or the instructor ends the scenario.

- Manual Scenario Mode**
- Scenario 1: SHOCK - SHOCK**
- Scenario 2: NO SHOCK - SHOCK**
- Scenario 3: SHOCK - NO SHOCK**
- Scenario 4: NO SHOCK - NO SHOCK**

Training Controls:



Pads attached/not attached to victim's bare chest (toggle-switch).



NO SHOCK advised on next heart analysis (manual mode only).



PUSH HARDER prompt (only when CPR Metronome is beeping).



GOOD COMPRESSIONS prompt (only when CPR Metronome is beeping).



Pause (resume) current scenario.



End current scenario.

ZOLL AED Plus® Trainer2:



Press to turn on Trainer2.
Press again to start current scenario (as selected using Remote Controller).
Press and hold down to turn off Trainer2.



Green light remains on = Trainer2 is on and ready for use.
Green light off = Trainer2 is turned off or has no battery power.
Red Light steady = Last command from Remote Controller accepted.
Red Light 3 flashes = Last command from Remote Controller is invalid.

1

SETUP and TEST

Trainer2 unit:

Action:

Install 4 "D" alkaline batteries in the AED Plus Trainer2.



Press **On/Off** Button to turn Trainer2 on.



Green light turns on, indicating Trainer2 is ready for use.

Remote Controller:

Action:



Install 2 "AA" batteries in handheld remote.
Turn the Remote Controller on by pushing the **View/Enter** Button.



Point Remote Controller at unit and press **View/Enter** Button again. Green light on Remote Controller will show Manual Scenario Mode is selected.

Manual Scenario Mode
View/Enter Button

Trainer2 unit:

Action:



Turn Trainer2 unit off (hold down **On/Off** Button).

Setup test is now complete. Trainer2 unit and Remote Controller are ready for use.

NOTE The Remote Controller lights indicate the current training settings while the Trainer2 simulates the actual AED Plus behavior.

2

SELECT SCENARIO SETTINGS

Trainer2 unit:

Action:



Press the **On/Off** Button to turn on the Trainer2.

Remote Controller:

Action:



Press the **View/Enter** Button once to turn on the Remote Controller.



Use **Move Up** Button to select desired scenario.



Use **CPR Metronome On/Off** Button to turn Real CPR Help® (beeping) on or off for this scenario.



Press the **View/Enter** Button once to transmit the selected Training Scenario to the Trainer2 unit.

NOTE The Remote Controller will turn itself off after 10 seconds. The Remote Controller will remember the last scenario setting which reappears the next time the **View/Enter** Button is pressed.

3

SCENARIO CONTROLS

Remote Controller: Action:

Analysis - VF



VF (Ventricular Fibrillation-shock advised) is the default simulated heart rhythm in Manual Scenario.

For the next analysis to result in **NO SHOCK ADVISED**, push the **No Shock** button.

NOTE Once the No Shock button has been pushed, the next simulated heart analysis will advise NO Shock. This setting cannot be reversed for the next analysis.



In all scenarios the **Pads** Button must be pressed once to proceed with analysis. If pressed again, it will begin to prompt for attaching pads.

3a

CPR Prompts

The Instructor uses the Remote Controller to control prompts during a Training Scenario's CPR period. The instructor can use the following buttons to simulate the use of an AED Plus defibrillator with Real CPR Help.

Remote Controller: Action:



Push Harder Button

Pressing this button during the CPR period causes the Trainer2 to issue the **PUSH HARDER** voice prompt. Instructors should press this button when a student's chest compressions are too shallow.



Good Compressions Button

Pressing this button during the CPR period causes the Trainer2 to issue the **GOOD COMPRESSIONS** voice prompt. Instructors should press this button to provide students with positive feedback when they respond to the **PUSH HARDER** prompt by improving the depth of chest compressions

NOTE Accurately imitating the student's AED Plus unit requires instructors be familiar with the protocol currently followed by the AED Plus for **PUSH HARDER** and **GOOD COMPRESSIONS** prompts.

Trainer2 unit:

Action:



The **Push Harder** and **Good Compressions** Buttons are only valid during the CPR interval. The red light on the Trainer2's status window emits one long flash if the prompt is currently valid.



If the **Push Harder** or **Good Compressions** Buttons are pushed when the Trainer2 is not beeping, the unit's red light emits three quick flashes, indicating that the prompt is currently invalid.

3b

PAUSE/END SCENARIO

Remote Controller: Action:



In all scenarios (manual and automatic), the rescue can be paused by pressing the **Pause/Resume** Button once. Press it a second time to resume the scenario.



In all scenarios (manual and automatic), the rescue can be ended by pressing the **End** Button once.

NOTE For further information about the functioning of the clinical AED Plus, see the AED Plus Operator's Guide and Administrator's Guide.

TROUBLESHOOTING

AED Plus Trainer2 problem/
display:

Trainer2 does not respond
(green light in status window
is not lit) when **On/Off**
Button is pushed



Remote Controller does not
respond (scenario
selection is not lit)
when **View/Enter**
Button is pushed.



Trainer2 does not beep or
react to Remote Controller
during CPR interval.

Trainer2 announces breathing
prompts, **OPEN AIRWAY** and
CHECK BREATHING.

With new batteries in both,
Trainer2 fails to respond to
Remote Controller.

Action:

Replace all four D batteries in
Trainer2 unit.

Replace both AA batteries in
Remote Controller.

Re-set scenario and make
certain that CPR metronome
light is on.

Unit is configured for ERC.
Change to AHA configuration.
(See Switching AHA/ERC
Configuration section of this
Operator's Guide).

Contact ZOLL
Technical Service.

MAINTENANCE

1. Keep Trainer2 and Remote Controller clean, undamaged, and free of excessive wear.
2. Verify that there are no cracks or loose parts in Trainer2 and Remote Controller.
3. Replace Training electrodes as necessary.

CLEANING

- Clean the Trainer2 unit and remote controller with soft, damp cloth using soap and water, then wipe dry immediately.
- Do not immerse any parts in water.
- Do not clean with ketones (e.g., MEK, acetone).
- Avoid using abrasives on any surface.
- Do not sterilize the Trainer2 or the Remote Controller.

ACCESSORIES

You can order the following Trainer2 accessories from the ZOLL Customer Service Department or your distributor:

Item:	REF
1. CPR-D padz® Training Electrodes (1 Set)	8900-0804-01
2. CPR-D padz Training Gel (5 Pair)	8900-0803-01
3. stat-padz® Training Electrodes (1 Set)	8900-0805-01
4. pedi-padz® II Training Electrodes (case of 6)	8900-000861-01
5. Trainer2 Remote Controller	8008-0007
6. Trainer Carrying Case	8000-0375-01
7. Trainer Demo Manikin	8000-0835-01

WARNINGS

- Product intended for indoor use only.
- This device is for training only. It will not deliver defibrillation energy.
- Do not attempt to plug AED Plus or AED Pro simulator into connector of Trainer2.
- Do not attempt to take Trainer2 or Remote Controller apart. Remove battery compartment doors only to replace batteries.
- Medical Electrical Equipment needs special precautions regarding EMC and needs to be installed and put into service according to the EMC information provided in accompanying documents.
- Portable and Mobile RF Communications equipment can affect Medical Electrical Equipment.

SCENARIOS

- 1 *Simulated heart rhythm:* *Student Response:*
VF 1 Shock
VF 2 CPR
VF 3 Shock
VF 4 CPR
Done
- 2 *Simulated heart rhythm:* *Student Response:*
ASYSTOLE 1 No Shock
ASYSTOLE 2 CPR
VF 1 Shock
VF 2 CPR
Done
- 3 *Simulated heart rhythm:* *Student Response:*
VF 1 Shock
VF 2 CPR
NSR 1 No Shock
NSR 2 CPR
Done
- 4 *Simulated heart rhythm:* *Student Response:*
ASYSTOLE 1 No Shock
ASYSTOLE 2 CPR
ASYSTOLE 3 No Shock
ASYSTOLE 4 CPR
Done

SWITCHING AHA/ERC CONFIGURATION

American Heart Association (AHA) configuration: No prompts for breathing.

European Resuscitation Council (ERC) configuration: Contains breathing prompts **OPEN AIRWAY** and **CHECK BREATHING**.

To switch between either AHA or ERC configuration, do the following:

- With the Trainer2 unit off, hold down the **Shock** Button. Press **On/Off** Button once. Green light will illuminate at status window to indicate AHA configuration (prompt sequence will contain no breathing prompts).
- Press **Shock** Button again. Green light will turn off and red light will illuminate at status window to indicate ERC configuration (prompt sequence will contain breathing).
- Press **Shock** Button again to revert to AHA configuration (green light on and red light off).
- When configured as desired (green = AHA, red = ERC), hold down **On/Off** Button to turn unit off to store configuration.

NOTE Additional Remote Controllers are available.
Contact ZOLL Medical Corporation or your distributor.

SYMBOLS

The following symbols appear on the ZOLL AED Trainer2:

Authorized representative in the European Community.

Return to a collection site intended for waste electrical and electronic equipment (WEEE). Do not dispose of in unsorted trash.

Manufacturer. Caution.

Temperature limitation. Catalogue Number.

ZOLL, AED Plus, Real CPR Help, CPR-D padz, stat-padz and pedi-padz are trademarks or registered trademarks of ZOLL Medical Corporation in the United States and/or other countries. Copyright 2012 by ZOLL Medical Corporation. All rights reserved.

Conformité Européenne - Complies with Low Voltage Directive 73/23/EEC.

AED Plus® Trainer2

Automated External Defibrillator Trainer2

Operator's Guide

AHA/ERC 2010

ZOLL

For technical support or repair:

ZOLL Medical Corporation
269 Mill Road
Chelmsford, MA USA 01824-4105

Attn: Technical Service Department

Tel: 978-421-9655 800-348-9011
Fax: 978-421-0001

ZOLL Medical Europe
Newtonweg 18
6662 PV ELST
The Netherlands

Tel: +31 (0) 481 366410
Fax: +31 (0) 481 366411

International Service: Contact your
local distributor.

© 2012 ZOLL Medical Corporation
TRAINING USE ONLY - NOT FOR CLINICAL USE
REF 9650-0090-01 Rev. C

(844) 277-2527 www.GoRescue.com